TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **POTASSIUM** MANUFACTURE DATE: 10/08/2009
CATALOG #: P605-50
LOT #: 19553 EXP. DATE: 04/2011

COMPONENTS:
1. POTASSIUM REAGENT LOT #: 19551 EXP. DATE: 07/2011
2. POTASSIUM STANDARD LOT #: 21024 EXP. DATE: 04/2011

BLANK: 0.053 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET (mEq/L)
15271	03/2012	3.7-4.7	4.3
15272	03/2012	4.9-6.9	5.9

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15261	3.9	4.0	4.0	2/2
15262	6.1	5.9	6.0	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 10/08/2009