TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **POTASSIUM** MANUFACTURE DATE: 05/10/2011
CATALOG #: P605-50
LOT #: 32409 EXP. DATE: 10/2013

COMPONENTS:					
1. POTASSIUM REAGENT	LOT #:	D110105022	EXP. DATE:	01/2014
2. POTASSIUM STANDARD	LOT #:	31800	EXP. DATE:	10/2013

BLANK: 0.053 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET (mEq/L)
15281	09/2012	3.4-5.2	4.3
15272	03/2012	4.9 – 6.9	5.9

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15281	4.3	4.1	4.2	2/2
15272	5.2	5.2	5.2	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 05/11/2011