TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **POTASSIUM** MANUFACTURE DATE: 08/12/2011
CATALOG #: P605-50
LOT #: 34095 EXP. DATE: 01/2014

COMPONENTS:	1. POTASSIUM REAGENT	LOT #:	D110501505	EXP. DATE:	05/2014
	2. POTASSIUM STANDARD	LOT #:	34079	EXP. DATE:	01/2014

BLANK: 0.054 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET (mEq/L)
15281	09/2012	3.4-5.2	4.3
15282	09/2012	4.9-7.4	6.2

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15281	4.1	4.6	4.4	2/2
15282	5.3	5.2	5.3	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 08/12/2011