



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **POTASSIUM** MANUFACTURE DATE: 11/18/2011
 CATALOG #: P605-50
 LOT #: 36126 EXP. DATE: 04/2014

COMPONENTS: 1. POTASSIUM REAGENT LOT #: D110506505 EXP. DATE: 05/2014
 2. POTASSIUM STANDARD LOT #: 35689 EXP. DATE: 04/2014

BLANK: 0.031 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15281	09/2012	3.1 – 4.6	3.8
15282	09/2012	5.2 – 7.8	6.5

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	3.9	3.9	3.9	2/2
15282	6.4	6.2	6.3	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Nour Qasqas
 Director of Quality Control

DATE: 11/18/2011