TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

# CERTIFICATE OF ANALYSIS
# (FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **POTASSIUM**
CATALOG #: P605-50
LOT #: 61667

MANUFACTURE DATE: 06/24/2015
EXP. DATE: 12/2017

COMPONENTS:
1. POTASSIUM REAGENT LOT #: 61584 EXP. DATE: 12/2017
2. POTASSIUM STANDARD LOT #: 61585 EXP. DATE: 12/2017

## TOLERANCE LEVELS OF SERUM CONTROLS

| SERA ID | EXP. DATE | TOLERANCE RANGE (on TC-Matrix) | TARGET (mmol/l) |
|---|---|---|---|
| 27621 | 12/2015 | 3.8 – 4.6 | 4.2 |
| 27622 | 12/2015 | 5.2 – 7.8 | 6.5 |

## TEST VALUES/RESULTS

| SERA ID | MEAN | PASS/FAIL |
|---|---|---|
| 27621 | 4.2 | PASS |
| 27622 | 6.8 | PASS |

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Jacqueline Sanchez
Quality Control

DATE: 06/25/2015