



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **SODIUM** MANUFACTURE DATE: 12/16/2009
 CATALOG #: S600-50G
 KIT LOT #: 22883 EXP. DATE: 5/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. SODIUM FILTRATE REAGENT	S600-A	21744	05/2012
2. SODIUM ACID REAGENT	S600-B	21745	05/2012
3. SODIUM COLOR REAGENT	S600-C	21746	05/2012
4. SODIUM STANDARD (150 mEq/L)	S600-D	21747	05/2012

BLANK: .021 AT 550 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15271	03/2012	103 – 127	115
15272	03/2012	114 – 144	129

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	118.8	115.6	119	2/2
15272	121.4	116.6	119	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


 Quality Control

DATE:12/16/2009