



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **SODIUM** MANUFACTURE DATE: 04/26/2010
 CATALOG #: S600-50
 KIT LOT #: 25338 EXP. DATE: 06/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. SODIUM FILTRATE REAGENT	S600-A	24371	09/2012
2. SODIUM ACID REAGENT	S600-B	24372	09/2012
3. SODIUM COLOR REAGENT	S600-C	23073	06/2012
4. SODIUM STANDARD (150 mEq/L)	S600-D	23074	06/2012

BLANK: 1.560 AT 550 NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15271	03/2012	103 – 127	115
15272	03/2012	114 – 144	129

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	112.9	110.8	111.9	2/2
15272	133.7	132.0	132.9	2/2

COMMENTS: VISUAL CHECK OK.

APPROVE D BY:	 Quality Control	DATE :	<u>04/26/2010</u>
------------------	--	-----------	-------------------