

TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
 Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
 Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **SODIUM** MANUFACTURE DATE: 6/10/2010
 CATALOG #: S600-50
 KIT LOT #: 26125 EXP. DATE: 9/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. SODIUM FILTRATE REAGENT	S600-A	25698	11/2012
2. SODIUM ACID REAGENT	S600-B	25699	11/2012
3. SODIUM COLOR REAGENT	S600-C	24369	9/2012
4. SODIUM STANDARD (150 mEq/L)	S600-D	23074	06/2012

BLANK: N/A AT 550 NM


TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15271	03/2012	103 – 127	115
15272	03/2012	114 – 144	129

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	108.9	110.0	109.5	2/2
15272	127.5	102.7	124.1	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:	 Quality Control	DATE:	6/11/10
-----------------	--	-------	---------