



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **SODIUM** MANUFACTURE DATE: 11/19/2010  
 CATALOG #: S600-50  
 KIT LOT #: 28982 EXP. DATE: 09/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT#:</u>	<u>EXP. DATE:</u>
1. SODIUM FILTRATE REAGENT	S600-A	28596	04/2013
2. SODIUM ACID REAGENT	S600-B	28597	04/2013
3. SODIUM COLOR REAGENT	S600-C	24369	09/2012
4. SODIUM STANDARD (150 mEq/L)	S600-D	26790	12/2012

BLANK: 1.458 AT 550 NM

**TOLERANCE LEVELS OF CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mEq/L)
15271	03/2012	103 – 127	115
15272	03/2012	114 – 144	129

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	108.9	104.5	106.7	2/2
15272	118.8	119.1	118.95	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:	<i>V. Casey</i> Quality Control	DATE:	11/19/2010
-----------------	------------------------------------	-------	------------