



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: TC SERUM CONTROL LEVEL II MANUFACTURE DATE: 08/13/2010

LOT #: 26858/15282 EXP. DATE: 09/2012

COMPONENT: TC SERUM CONTROL LEVEL II LOT#: 15282 EXP. DATE: 09/2012

REAGENT (ALT) BLANK VS WATER: 0.043 AT 340 NM


TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> <u>CHOLESTEROL</u>	<u>TARGET</u>
15282	09/2012	91-113.4	102

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15282	106.1	94.1	100.1	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: 
 Quality Control
 QA/QC

DATE: 08/13/2010