



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: TC SERUM CONTROL LEVEL I MANUFACTURE DATE: 7/24/2009

LOT #: 15271 EXP. DATE: 03/2012

COMPONENT: TC SERUM CONTROL LEVEL I LOT#: 15271 EXP. DATE: 3/2012

REAGENT (INORGANIC PHOSPHORUS) BLANK VS WATER: 0.036 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> <u>INORGANIC PHOSPH.</u>	<u>TARGET</u>
15271	03/2012	3.0 – 4.4	3.7
15272	03/2012	5.5 – 7.9	6.7

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	3.4	3.9	3.7	2/2
15272	6.4	5.8	6.1	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:



DATE: 9/23/2009

QA/QC