



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: TC SERUM CONTROL LEVEL I MANUFACTURE DATE: 08/13/2010

LOT #: 26859/15281 EXP. DATE: 09/2012

COMPONENT: TC SERUM CONTROL LEVEL I LOT#: 15281 EXP. DATE: 09/2012

REAGENT (CREATININE) BLANK VS WATER: 0.043 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> <u>CHOLESTEROL</u>	<u>TARGET</u>
15281	09/2012	207.5-302.0	234.5

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	239.6	238.2	238.9	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 08/13/2010
 Quality Control :
 QA/QC