



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **TOTAL PROTEIN** MANUFACTURE DATE: 01/13/2011
 CATALOG #: T528-480
 LOT #: 30096 EXP. DATE: 04/2013

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
1. TOTAL PROTEIN REAGENT	T528-A	M1100501016	05/2013
2. TOTAL PROTEIN STANDARD (5.0 g/dL)	T528-B	25220	04/2013

BLANK: 0.093 AT 550 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (g/dL)
15271	03/2012	4.8 – 7.0	5.9
15272	03/2012	3.2 – 4.8	4.0

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	6.7	6.7	6.7	2/2
15272	4.1	3.6	3.85	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 01/13/2011
Quality Control