



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME:	TOTAL PROTEIN	MANUFACTURE DATE:	06/15/2015
CATALOG #:	T528-480	EXP. DATE:	09/2017
LOT #:	61453		

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
1. TOTAL PROTEIN REAGENT	T528-A	59113	01/2018
2. TOTAL PROTEIN STANDARD (5.0 g/dL)	T528-B	57373	09/2017

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC Matrix)	<u>TARGET</u> (g/dL)
27621	12/2015	6.0 – 7.4	6.7
27622	12/2015	3.9 – 4.8	4.3

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	7.1	PASS
27622	4.7	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez
 Quality Control

DATE: