



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **TOTAL PROTEIN** MANUFACTURE DATE: 03/08/2010
 CATALOG #: T528-480
 LOT #: 24527 EXP. DATE: 08/2012

<u>COMPONENTS:</u>	<u>CATALOG #:</u>	<u>LOT #:</u>	<u>EXP. DATE:</u>
1. TOTAL PROTEIN REAGENT	T528-A	23416	01/2013
2. TOTAL PROTEIN STANDARD (5.0 g/dL)	T528-B	20406	08/2013

BLANK: 0.69 AT 550 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (g/dL)
15271	03/2012	4.7 – 7.0	5.9
15272	03/2012	3.2 – 4.9	4.0

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	6.2	6.2	6.2	2/2
15272	3.4	3.4	3.4	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Basen* DATE: 03/08/2010
 Quality Control