



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **TRIGLYCERIDE-GPO (LIQUID)** MANUFACTURE DATE: 05/27/2010
 CATALOG #: T532-480
 LOT #: 25664 EXP. DATE: 11/2011

COMPONENTS: 1. REAGENT LOT#: 25675 EXP. DATE: 11/2011
 2. STANDARD LOT#: 23965 EXP. DATE: 12/2012

BLANK: 0.029 AT 550NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15271	03/2012	121-203	162
15272	03/2012	57-95	76

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	172.0	189.0	180.5	2/2
15272	79.2	71.8	75.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Gascón* DATE: 05/27/2010
 Quality Control