

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.  
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880  
Fax: (714) 463-1169  
[www.tecodiag.com](http://www.tecodiag.com)

**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **TRIGLYCERIDE-GPO (LIQUID)** MANUFACTURE DATE: 11/10/2011  
CATALOG #: T532-480  
LOT #: 35794 EXP. DATE: 04/2013

COMPONENTS: 1. REAGENT LOT#: 35793 EXP. DATE: 04/2013  
2. STANDARD LOT#: 32489 EXP. DATE: 06/2014

BLANK: 0.006 AT 550NM

**TOLERANCE LEVELS OF CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dL)</u>
15281	09/2012	156 - 234	207
15282	09/2012	65 - 113	89

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	206	207	207	2/2
15282	83	83	83	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 11/10/2011

Nour Qasqas  
Quality Control Lab. Director