



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **TRIGLYCERIDE – GPO**      MANUFACTURE DATE: 08/6/2009  
 CATALOG #: T531-150  
 LOT #: 20344      EXP. DATE: 04/2012

COMPONENTS:    1. REAGENT                      LOT#: 18961    EXP. DATE: 04/2012  
                          2. STANDARD                      LOT#: 20199    EXP. DATE: 07/2012

BLANK:            0.022            AT 550 NM

STANDARD:      200 mg/dL

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET (mg/dL)</u>
15261	08/2011	148 – 200	174
15262	08/2011	67 – 90	79

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15261	192.6	190.9	191.8	2/2
15262	89.0	71.6	80.3	2/2

COMMENT: VISUAL CHECK OK.

APPROVED  
BY:

Quality Control

DATE: 8/6/2009