



**CERTIFICATE OF ANALYSIS**  
**(FINISHED KIT QUALITY CONTROL)**

PRODUCT NAME: **TRIGLYCERIDE – GPO**      MANUFACTURE DATE: 11/24/2009  
 CATALOG #: T531-400  
 LOT #: 22545      EXP. DATE: 07/2012

COMPONENTS: 1. REAGENT      LOT#: 20321      EXP. DATE: 08/2012  
 2. STANDARD      LOT#: 20199      EXP. DATE: 07/2012

BLANK: 0.034      AT 550 NM

STANDARD: 200 mg/dL

**TOLERANCE LEVELS OF SERUM CONTROLS**

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dL)
15271	03/2012	137 – 207	174
15272	03/2012	59 – 95	79

**TEST VALUES/RESULTS**

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15271	166.3	159.2	162.75	2/2
15272	66.8	72.9	69.85	2/2

COMMENT: VISUAL CHECK OK.

APPROVED BY: *V. [Signature]*      DATE: 11/24/2009  
 Quality Control