TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME:	**TRIGLYCERIDE-GPO (LIQUID)**	MANUFACTURE DATE:	12/06/2010
CATALOG #:	T532-480		
LOT #:	29231	EXP.DATE:	04/2012

COMPONENTS:	1. REAGENT	LOT#:	29230	EXP. DATE:	04/2012
	2. STANDARD	LOT#:	23965	EXP. DATE:	12/2012

BLANK: 0.009 AT 550NM

TOLERANCE LEVELS OF CONTROLS

SERA ID	EXP. DATE	TOLERANCE RANGE	TARGET (mg/dL)
15271	03/2012	121-203	162
15272	03/2012	57-95	76

TEST VALUES/RESULTS

SERA ID	1	2	MEAN	# PASS
15271	187.3	175.3	177.3	2/2
15272	76.3	71.1	73.7	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 12/06/2010