



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **TRIGLYCERIDE-GPO (LIQUID)** MANUFACTURE DATE: 06/28/2011
CATALOG #: T532-480
LOT #: 33353 EXP. DATE: 12/2012

COMPONENTS: 1. REAGENT LOT#: 33009 EXP. DATE: 12/2012
2. STANDARD LOT#: 32489 EXP. DATE: 06/2014

BLANK: 0.026 AT 550NM

TOLERANCE LEVELS OF CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dL)</u>
15281	09/2012	125 – 188	157
15272	03/2012	57-95	76

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	182.7	179.9	181.3	2/2
15272	61.6	73.4	67.5	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: 
Quality Control DATE: 06/28/2011