

TECO DIAGNOSTICS



1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: TRIGLYCERIDE-GPO (LIQUID) MANUFACTURE DATE: 08/28/2014
CATALOG #: T532-480
LOT #: 56895 EXP. DATE: 11/2015

COMPONENTS: REAGENT LOT#: 55124 EXP. DATE: 11/2015
STANDARD LOT#: 56521 EXP. DATE: 08/2017

TOLERANCE LEVELS OF CONTROLS


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (on TC-Matrix)	<u>TARGET</u> (mg/dL)
27611	05/2015	136.8 – 205.2	171.0
27602	03/2015	50.6 – 84.3	67.4

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	168.4	PASS
27602	66.3	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY: _____


Jacqueline Sanchez
Quality Control

DATE: _____

