



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URIC ACID** MANUFACTURE DATE: 04/19/2010
 CATALOG #: U580-240
 LOT #: 25128 EXP. DATE: 12/2012

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT #</u>	<u>EXP. DATE</u>
1. URIC ACID REAGENT	U580-A	23433	12/2012
2. URIC ACID STANDARD (5.0 mg/dl)	U580-B	23434	12/2012

BLANK: 0.008 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> <u>(mg/dl)</u>
15271	03/2012	3.9-6.1	5.0
15272	03/2012	6.6-10.0	8.3

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15261	5.1	4.9	5.0	2/2
15262	8.0	8.4	8.2	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Casey* DATE: 04/19/2010
 Quality Control