



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URIC ACID** MANUFACTURE DATE: 03/07/2011
 CATALOG #: U580-240
 LOT #: 30439 EXP. DATE: 02/2014

<u>COMPONENTS</u>	<u>CATALOG #</u>	<u>LOT #</u>	<u>EXP. DATE</u>
1. URIC ACID REAGENT	U580-A	30082	02/2014
2. URIC ACID STANDARD (5.0 mg/dl)	U580-B	30823	03/2014

BLANK: 0.022 AT 505 NM

TOLERANCE LEVELS OF SERUM CONTROLS

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u>	<u>TARGET</u> (mg/dl)
15281	09/2012	3.5-5.3	4.4
15272	03/2012	6.6 - 10.0	8.3

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>1</u>	<u>2</u>	<u>MEAN</u>	<u># PASS</u>
15281	5.0	5.0	5.0	2/2
15272	8.1	9.0	8.6	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: *V. Casey* DATE: 03/07/2011
 Quality Control