



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URIC ACID**
 CATALOG #: U580-240
 LOT #: 57942

MANUFACTURE DATE: 11/05/2014
 EXP. DATE: 01/2017

COMPONENTS: 1. URIC ACID REAGENT
 2. URIC ACID STANDARD

LOT #: 54188 EXP. DATE: 04/2017
 LOT #: 52744 EXP. DATE: 01/2017

TOLERANCE LEVELS OF SERUM CONTROL


<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
27611	05/2015	4.7 – 6.6	5.6
27602	03/2015	7.3 – 10.3	8.8

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27611	5.8	PASS
27602	10.0	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:


 Jacqueline Sanchez
 Quality Control

DATE:

11/06/2014