



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URIC ACID** MANUFACTURE DATE: 02/18/2015
 CATALOG #: U580-240
 LOT #: 59371 EXP. DATE: 10/2017

COMPONENTS: 1. URIC ACID REAGENT LOT #: 59112 EXP. DATE: 02/2018
 2. URIC ACID STANDARD LOT #: 57526 EXP. DATE: 10/2017

TOLERANCE LEVELS OF SERUM CONTROL

<u>SERA ID</u>	<u>EXP. DATE</u>	<u>TOLERANCE RANGE</u> (On TC-84)	<u>TARGET</u> (mg/dl)
27621	05/2015	5.2 – 7.3	6.3
27602	03/2015	7.9 – 11.1	9.5

TEST VALUES/RESULTS

<u>SERA ID</u>	<u>MEAN</u>	<u>PASS/FAIL</u>
27621	5.8	PASS
27602	9.1	PASS

COMMENTS: VISUAL CHECK OK.

APPROVED BY:

Jacqueline Sanchez
 Quality Control

DATE: 02/18/2015