


REF : 3734-3735-3736-3752-3753-3754-3820-3821-3822
 LOT : 188
 : 2009, August 19th

Coulter Counter Models (with III Diff. option; for ActT see separate table)

MELET MS9 NOTE : Please refer to next page for Erma Values

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.4	7.7	0.8	19.8	2.0
RBC 10 ¹² /l	2.34	0.12	4.65	0.15	5.54	0.18
Hb g/dl	6.1	0.4	13.5	0.5	17.7	0.6
Hb mmol/l	3.8	0.2	8.4	0.3	11.0	0.4
Hct %	17.6	1.5	39.1	2.4	51.0	3.0
Hct l/l	0.176	0.015	0.391	0.024	0.510	0.030
MCV fl	75	4	84	4	92	4
MCH pg	26.1	2.4	29.0	2.8	31.9	2.8
MCH fmol	1.62	0.14	1.80	0.11	1.98	0.11
MCHC g/dl	34.8	3.0	34.6	3.0	34.7	3.0
MCHC mmol/l	21.6	1.8	21.4	1.8	21.6	1.8
RDW %	15.0	3.0	14.6	3.0	13.7	3.0
Plts 10 ⁹ /l	73	15	245	30	480	55
MPV fl	7.4	2.0	8.0	1.5	7.6	1.5
PDW %	17.7	2.5	16.7	2.5	16.5	2.5
LYMP %	59.0	5.0	33.5	4.0	15.5	4.0
MONO %	11.0	4.0	6.0	4.0	4.5	3.0
GRAN %	30.0	4.0	60.5	5.0	80.0	5.0
LYMP #	1.3	0.2	2.6	0.4	3.1	0.8
MONO #	0.2	0.1	0.5	0.4	0.9	0.6
GRAN #	0.7	0.1	4.7	0.4	15.8	1.0

ABX Roche: Micros; Bayer: Advia 60

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.3	0.4	7.3	0.8	18.3	2.0
RBC 10 ¹² /l	2.35	0.12	4.61	0.15	5.53	0.18
Hb g/dl	5.8	0.4	13.5	0.5	17.6	0.6
Hb mmol/l	3.6	0.2	8.4	0.3	10.9	0.4
Hct %	16.3	1.5	37.6	2.4	50.4	3.0
Hct l/l	0.163	0.015	0.376	0.024	0.504	0.030
MCV fl	70	4	82	4	91	4
MCH pg	24.7	2.4	29.2	2.8	31.9	2.8
MCH fmol	1.53	0.14	1.81	0.11	1.98	0.11
MCHC g/dl	35.4	3.0	35.8	3.0	35.0	3.0
MCHC mmol/l	22.0	1.8	22.2	1.8	21.7	1.8
RDW %	13.7	3.0	15.0	3.0	14.1	3.0
Plts 10 ⁹ /l	80	15	243	30	449	55
MPV fl	7.1	2.0	7.6	1.5	7.2	1.5
LYMP %	51.0	5.0	29.7	4.0	14.1	4.0
MONO %	10.5	4.0	5.9	4.0	3.5	3.0
GRAN %	38.5	4.0	64.4	5.0	82.3	5.0
LYMP #	1.2	0.3	2.2	0.4	2.6	0.8
MONO #	0.2	0.1	0.4	0.3	0.6	0.6
GRAN #	0.9	0.2	4.7	0.6	15.1	1.2

Abbott Cell-Dyn: 1300, 1500, 1600, 1700, 2000 series

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.4	0.4	7.4	0.8	17.7	2.0
RBC 10 ¹² /l	2.37	0.12	4.61	0.15	5.43	0.18
Hb g/dl	6.0	0.4	13.3	0.5	17.5	0.6
Hb mmol/l	3.7	0.2	8.3	0.3	10.9	0.4
Hct %	17.8	1.5	39.2	2.4	50.5	3.0
Hct l/l	0.178	0.015	0.392	0.024	0.505	0.030
MCV fl	75	4	85	4	93	4
MCH pg	25.3	2.4	28.9	2.8	32.2	2.8
MCH fmol	1.57	0.14	1.79	0.11	2.00	0.11
MCHC g/dl	33.8	3.0	33.9	3.0	34.7	3.0
MCHC mmol/l	20.9	1.8	21.1	1.8	21.5	1.8
Plts 10 ⁹ /l	69	15	240	30	467	55
MPV fl	8.3	2	8.5	1.5	7.8	1.5
LYMP %	59.0	5.0	33.5	4.0	15.5	4.0
MID %	11.0	4.0	6.0	4.0	4.5	3.0
GRAN %	30.0	4.0	60.5	5.0	80.0	5.0
LYMP #	1.4	0.3	2.5	0.4	2.7	0.8
MID #	0.3	0.1	0.4	0.3	0.8	0.6
GRAN #	0.7	0.2	4.5	0.6	14.2	1.2

Coulter Counter: Act diff, Act diff 2

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.4	7.7	0.8	19.8	2.0
RBC 10 ¹² /l	2.34	0.12	4.65	0.15	5.54	0.18
Hb g/dl	6.1	0.4	13.5	0.5	17.7	0.6
Hb mmol/l	3.8	0.3	8.4	0.3	11.0	0.4
Hct %	17.6	1.5	39.1	2.4	51.0	3.0
Hct l/l	0.176	0.015	0.391	0.024	0.510	0.030
MCV fl	75	4	84	4	92	4
MCH pg	26.1	2.4	29.0	2.8	31.9	2.8
MCH fmol	1.62	0.10	1.80	0.11	1.98	0.12
MCHC g/dl	34.8	3.0	34.6	3.0	34.7	3.0
MCHC mmol/l	21.6	1.8	21.4	1.8	21.6	1.8
RDW %	15.0	3.0	14.6	3.0	13.7	3.0
Plts 10 ⁹ /l	73	15	245	30	480	55
MPV fl	7.4	2.0	8.0	1.5	7.6	1.5
LYMP %	62.0	6.0	32.5	4.0	14.5	4.0
MONO %	8.0	5.0	4.5	4.0	3.5	3.0
GRAN %	30.0	4.0	63.0	5.0	82.0	5.0
LYMP #	1.4	0.2	2.5	0.4	2.9	0.8
MONO #	0.2	0.1	0.3	0.3	0.7	0.6
GRAN #	0.7	0.1	4.9	0.4	16.2	1.0

Coulter Counter: Act 8, Act 10

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
LYMP %	47.5	6.0	24.5	4.0	10.0	4.0
LYMP #	1.0	0.2	1.9	0.4	2.0	0.8

Please refer to Table Act Diff for CBC parameters


Nihon Kohden: Celltac and MEK series (not for 5-part diff)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.4	7.4	0.8	18.2	2.0
RBC 10 ¹² /l	2.32	0.12	4.61	0.15	5.46	0.18
Hb g/dl	6.0	0.4	13.8	0.5	18.9	0.6
Hb mmol/l	3.7	0.2	8.6	0.3	11.7	0.4
Hct %	17.9	1.5	39.3	2.4	52.8	3.0
Hct l/l	0.179	0.015	0.393	0.024	0.528	0.030
MCV fl	77	4	85	4	97	4
MCH pg	25.9	2.4	29.9	2.8	34.6	2.8
MCH fmol	1.60	0.14	1.86	0.11	2.15	0.11
MCHC g/dl	33.6	3.0	35.0	3.0	35.8	3.0
MCHC mmol/l	20.8	1.8	21.7	1.8	22.2	1.8
RDW %	14.0	3.0	14.1	3.0	13.0	3.0
Plts 10 ⁹ /l	75	15	250	30	469	55
MPV fl	7.4	2	7.0	1.5	6.2	1.5
LYMP %	62.9	5.0	35.6	4.0	18.7	4.0
MONO %	12.0	4.0	6.5	4.0	4.3	3.0
GRAN %	25.1	4.0	57.9	5.0	77.0	5.0
LYMP #	1.4	0.3	2.6	0.4	3.4	0.8
MONO #	0.3	0.1	0.5	0.3	0.8	0.6
GRAN #	0.6	0.2	4.3	0.6	14.0	1.2

Danam / Drew Excel18 = MWI BT2100

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.3	0.4	7.8	0.8	20.0	2.0
RBC 10 ¹² /l	2.46	0.12	4.57	0.15	5.42	0.18
Hb g/dl	6.1	0.4	13.4	0.5	17.8	0.6
Hb mmol/l	3.8	0.2	8.3	0.3	11.0	0.4
Hct %	18.3	1.5	38.8	2.4	50.6	3.0
Hct l/l	0.183	0.015	0.388	0.024	0.506	0.030
MCV fl	75	4	85	4	93	4
MCH pg	25.0	2.4	29.3	2.8	32.8	2.8
MCH fmol	1.55	0.14	1.82	0.11	2.03	0.11
MCHC g/dl	33.5	3.0	34.5	3.0	35.1	3.0
MCHC mmol/l	20.8	1.8	21.4	1.8	21.8	1.8
RDW %	11.3	3.0	11.0	3.0	10.9	3.0
Plts 10 ⁹ /l	84	15	259	30	467	55
MPV fl	6.2	2	7.2	1.5	6.8	1.5
LYMP %	51.7	5.0	26.9	4.0	11.6	4.0
MONO %	11.4	4.0	5.5	4.0	3.5	3.0
GRAN %	36.9	4.0	67.5	5.0	84.9	5.0
LYMP #	1.2	0.3	2.1	0.4	2.3	0.8
MONO #	0.3	0.1	0.4	0.3	0.7	0.6
GRAN #	0.8	0.2	5.2	0.6	17.0	1.2

LOT : 188

 : 2009, August 19th

Diatron: Abacus, Arcus **

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.4	7.4	0.8	19.4	2.0
RBC 10 ¹² /l	2.48	0.12	4.82	0.15	5.69	0.20
Hb g/dl	5.9	0.4	13.2	0.5	17.1	0.6
Hb mmol/l	3.7	0.3	8.2	0.4	10.6	0.4
Hct %	19.8	1.5	43.1	2.4	55.8	3.0
Hct l/l	0.198	0.015	0.431	0.024	0.558	0.030
MCV fl	80	4	89	4	98	4
MCH pg	23.8	2.4	27.4	2.8	30.0	2.8
MCH fmol	1.48	0.14	1.70	0.11	1.86	0.11
MCHC g/dl	29.8	3.0	30.6	3.0	30.7	3.0
MCHC mmol/l	18.5	1.8	19.0	1.8	19.0	1.8
RDW-SD fl	52.7	3.0	57.5	3.0	57.5	3.0
RDW-CV %	16.8	3.0	16.4	3.0	14.9	3.0
Plts 10 ⁹ /l	69	15	271	30	512	55
MPV fl	7.2	1.5	8.0	1.5	7.8	1.5
PCT %	0.05	0.02	0.22	0.03	0.40	0.04
PCT ml/l	0.50	0.2	2.17	0.3	3.99	0.4
PDW-SD fl	10.1	2.5	10.7	2.5	9.9	3.0
PDW-CV %	39.0	2.5	37.6	2.5	35.6	3.0
LYMP %	56.1	5.0	32.2	4.0	13.1	4.0
MONO %	11.1	4.0	5.3	4.0	4.7	4.0
GRAN %	32.8	4.0	62.5	5.0	82.2	5.0
LYMP #	1.1	0.3	2.4	0.4	2.5	0.8
MONO #	0.2	0.1	0.4	0.3	0.9	0.6
GRAN #	0.7	0.2	4.6	0.6	15.9	1.2

** JT Baker reagents CyMet Abacus + Diluid Abacus = Use Reagent Code 2 in service settings!!!

AVL (Contraves): AVL - 816

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	63.8	5.0	39.9	4.0	18.4	4.0
MONO %	10.1	4.0	6.8	4.0	5.8	3.0
GRAN %	26.1	4.0	53.3	5.0	75.8	5.0
LYMP #	1.4	0.3	3.1	0.4	3.6	0.8
MONO #	0.2	0.1	0.5	0.3	1.1	0.6
GRAN #	0.6	0.2	4.1	0.6	15.0	1.2

AVL (Contraves): AVL - 818

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	71.6	5.0	42.0	4.0	19.4	4.0
MONO %	8.9	4.0	6.4	4.0	5.5	3.0
GRAN %	19.5	4.0	51.6	5.0	75.1	5.0
LYMP #	1.6	0.3	3.2	0.4	3.8	0.8
MONO #	0.2	0.1	0.5	0.2	1.1	0.4
GRAN #	0.4	0.2	4.0	0.6	14.9	1.2

ERMA PCE-170, ERMAX-18, PCE-210

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.4	7.9	0.8	20.7	2.0
RBC 10 ¹² /l	2.45	0.12	4.66	0.15	5.39	0.18
Hb g/dl	5.9	0.4	13.0	0.5	17.0	0.6
Hb mmol/l	3.7	0.2	8.1	0.3	10.6	0.4
Hct %	19.1	1.5	41.0	2.4	53.9	3.0
Hct l/l	0.191	0.015	0.410	0.024	0.539	0.030
MCV fl	78	4	88	4	100	4
MCH pg	24.1	2.4	27.9	2.8	31.5	2.8
MCH fmol	1.49	0.14	1.73	0.11	1.96	0.11
MCHC g/dl	30.9	3.0	31.7	3.0	31.5	3.0
MCHC mmol/l	19.2	1.8	19.7	1.8	19.6	1.8
RDW %	18.0	3.0	16.1	3.0	15.1	3.0
Plts 10 ⁹ /l	83	15	259	30	506	55
MPV fl	8.2	2.0	8.4	1.5	8.1	1.5
PDW %	16.5	2.5	15.3	2.5	14.1	2.5
LYMP %	51.0	5.0	35.3	4.0	27.1	4.0
MONO %	4.6	4.6	2.3	2.3	1.5	1.5
GRAN %	44.4	4.0	62.4	5.0	71.4	5.0
LYMP #	1.1	0.3	2.8	0.4	5.6	0.8
MONO #	0.1	0.1	0.2	0.2	0.3	0.3
GRAN #	1.0	0.2	4.9	0.6	14.8	1.2

Orphee MYTHIC 18

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.4	7.7	0.8	19.8	2.0
RBC 10 ¹² /l	2.34	0.12	4.65	0.15	5.54	0.18
Hb g/dl	6.1	0.4	13.5	0.5	17.7	0.6
Hb mmol/l	3.8	0.2	8.4	0.3	11.0	0.4
Hct %	19.0	1.5	41.4	2.4	53.2	3.0
Hct l/l	0.190	0.015	0.414	0.024	0.532	0.030
MCV fl	81	4	89	4	96	4
MCH pg	26.1	2.4	29.0	2.8	31.9	2.8
MCH fmol	1.62	0.14	1.80	0.11	1.98	0.11
MCHC g/dl	32.2	3.0	32.6	3.0	33.3	3.0
MCHC mmol/l	20.0	1.8	20.2	1.8	20.7	1.8
RDW %	15.0	3.0	14.6	3.0	14.7	3.0
Plts 10 ⁹ /l	73	15	239	30	468	55
MPV fl	8.1	2.0	8.4	1.5	7.8	1.5
LYMP %	59.3	5.0	32.3	4.0	15.1	4.0
MONO %	6.4	6.4	2.9	2.9	2.5	2.5
GRAN %	34.3	4.0	64.8	5.0	82.4	5.0
LYMP #	1.3	0.3	2.5	0.4	3.0	0.8
MONO #	0.1	0.1	0.2	0.2	0.5	0.5
GRAN #	0.8	0.2	5.0	0.6	16.3	1.2



Mallinckrodt Baker B.V.
P.O. Box 1 - 7400AA Deventer - Holland
Tel.: +31 (0)570 687500


The devices as mentioned in this insert comply with the
In Vitro Diagnostic Medical Devices Directive 98/79/EC.



8°C



LOT : 188

 : 2009, August 19th

Mindray BC-3200, BC-3000CT (new models)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	7.6	1.0	19.8	2.5
RBC 10 ¹² /l	2.40	0.18	4.75	0.24	5.75	0.30
Hb g/dl	6.2	0.4	13.7	0.6	18.4	0.8
Hb mmol/l	3.8	0.3	8.5	0.4	11.4	0.5
Hct %	18.2	1.5	40.6	2.0	53.5	2.4
Hct l/l	0.182	0.015	0.406	0.020	0.535	0.024
MCV fl	76.0	5.0	85.5	5.0	93.0	5.0
MCH pg	25.8	2.5	28.8	2.5	32.0	2.5
MCH fmol	1.60	0.16	1.79	0.16	1.99	0.16
MCHC g/dl	34.0	3.0	33.7	3.0	34.4	3.0
MCHC mmol/l	21.1	1.8	20.9	1.8	21.4	1.8
RDW-CV	14.0	3.0	14.0	3.0	13.0	3.0
RDW-SD	38.0	6.0	41.0	6.0	43.0	8.0
Plts 10 ⁹ /l	80	20	245	40	460	60
MPV fl	11.4	3.0	10.5	3.0	9.8	3.0
LYMP %	50.5	8.0	27.0	7.0	14.0	6.0
MONO %	16.5	6.0	8.5	6.0	5.5	5.5
GRAN %	33.0	7.0	64.5	7.0	80.5	8.0
LYMP #	1.0	0.2	2.1	0.6	2.8	1.2
MONO #	0.3	0.2	0.6	0.5	1.1	1.1
GRAN #	0.7	0.2	4.9	0.6	15.9	1.6

Mindray BC-2800, BC-2600 (software version lower than 2.0)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	1.8	0.5	7.6	1.0	20.7	2.5
RBC 10 ¹² /l	2.26	0.18	4.52	0.24	5.50	0.30
Hb g/dl	6.3	0.4	13.8	0.6	18.6	0.8
Hb mmol/l	3.9	0.3	8.6	0.4	11.5	0.5
Hct %	17.4	1.5	38.9	2.0	51.4	2.4
Hct l/l	0.174	0.015	0.389	0.020	0.514	0.024
MCV fl	77.0	5.0	86.0	5.0	93.5	5.0
MCH pg	27.9	2.5	30.5	2.5	33.8	2.5
MCH fmol	1.73	0.16	1.89	0.16	2.10	0.16
MCHC g/dl	36.2	3.0	35.5	3.0	36.2	3.0
MCHC mmol/l	22.5	1.8	22.0	1.8	22.4	1.8
RDW-CV	14.0	3.0	14.0	3.0	14.0	3.0
RDW-SD	37.0	6.0	41.0	6.0	43.0	8.0
Plts 10 ⁹ /l	65	20	245	40	500	60
MPV fl	8.7	3.0	9.2	3.0	8.7	3.0
PCT %	0.06	0.05	0.23	0.10	0.44	0.20
PCT ml/l	0.57	0.50	2.25	1.00	4.35	2.00
PDW fl	15.7	3.0	16.1	3.0	16.0	3.0
LYMP %	63.5	9.0	31.5	7.0	14.5	6.0
MONO %	7.5	7.0	5.0	5.0	6.0	5.0
GRAN %	29.0	7.0	63.5	7.0	79.5	8.0
LYMP #	1.1	0.2	2.4	0.6	3.0	1.3
MONO #	0.1	0.1	0.4	0.4	1.2	1.0
GRAN #	0.5	0.2	4.8	0.6	16.5	1.7

Mindray BC-2800, BC-2600 (software version 2.0 or higher)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	7.5	1.0	20.1	2.5
RBC 10 ¹² /l	2.31	0.18	4.62	0.24	5.56	0.30
Hb g/dl	6.3	0.4	13.7	0.6	18.3	0.8
Hb mmol/l	3.9	0.3	8.5	0.4	11.4	0.5
Hct %	17.3	1.5	39.3	2.0	52.0	2.4
Hct l/l	0.173	0.015	0.393	0.020	0.520	0.024
MCV fl	75.0	5.0	85.0	5.0	93.5	5.0
MCH pg	27.3	2.5	29.7	2.5	32.9	2.5
MCH fmol	1.69	0.16	1.84	0.16	2.04	0.16
MCHC g/dl	36.4	3.0	34.9	3.0	35.2	3.0
MCHC mmol/l	22.6	1.8	21.7	1.8	21.8	1.8
RDW-CV	16.0	3.0	15.5	3.0	14.5	3.0
RDW-SD	41.0	6.0	44.0	6.0	45.5	8.0
Plts 10 ⁹ /l	60	20	243	40	451	60
MPV fl	9.5	3.0	9.5	3.0	9.2	3.0
PCT %	0.06	0.05	0.23	0.10	0.41	0.20
PCT ml/l	0.57	0.50	2.31	1.00	4.15	2.00
PDW fl	14.8	3.0	14.8	3.0	14.7	3.0
LYMP %	64.5	9.0	32.5	7.0	15.0	6.0
MONO %	7.0	6.0	5.5	5.0	6.5	5.0
GRAN %	28.5	8.0	62.0	8.0	78.5	8.0
LYMP #	1.3	0.2	2.4	0.6	3.0	1.2
MONO #	0.1	0.1	0.4	0.4	1.3	1.0
GRAN #	0.6	0.2	4.7	0.7	15.8	1.7

Mindray BC-3000 (new models s/n A75xxx) , BC-2900, BC-1800

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.1	0.5	7.7	1.0	19.5	2.5
RBC 10 ¹² /l	2.42	0.18	4.83	0.24	5.78	0.30
Hb g/dl	6.3	0.4	13.9	0.6	18.4	0.8
Hb mmol/l	3.9	0.3	8.6	0.4	11.4	0.5
Hct %	18.4	1.5	41.5	2.0	54.3	2.4
Hct l/l	0.184	0.015	0.415	0.020	0.543	0.024
MCV fl	76.0	5.0	86.0	5.0	94.0	5.0
MCH pg	26.0	2.5	28.8	2.5	31.8	2.5
MCH fmol	1.62	0.16	1.79	0.16	1.98	0.16
MCHC g/dl	34.3	3.0	33.5	3.0	33.9	3.0
MCHC mmol/l	21.3	1.8	20.8	1.8	21.0	1.8
RDW-CV	14.0	3.0	14.0	3.0	13.0	3.0
RDW-SD	40.0	6.0	44.0	6.0	46.0	8.0
Plts 10 ⁹ /l	68	20	238	40	460	60
MPV fl	10.2	3.0	10.6	3.0	10.1	3.0
PCT %	0.07	0.05	0.25	0.10	0.46	0.20
PCT ml/l	0.69	0.50	2.52	1.00	4.65	2.00
PDW fl	16.1	3.0	15.7	3.0	15.6	3.0
LYMP %	50.0	8.0	26.5	7.0	15.0	6.0
MONO %	17.5	6.0	9.0	6.0	4.5	4.5
GRAN %	32.5	7.0	64.5	7.0	80.5	8.0
LYMP #	1.1	0.2	2.0	0.6	2.9	1.2
MONO #	0.4	0.2	0.7	0.5	0.9	0.9
GRAN #	0.7	0.2	5.0	0.6	15.7	1.6

Mindray: BC-3000 with Cymet Mindray CN-free (old models s/n A59xxx)

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.2	0.5	7.7	1.0	18.6	2.5
RBC 10 ¹² /l	2.19	0.18	4.48	0.24	5.48	0.30
Hb g/dl **	6.3	0.4	14.4	0.6	19.1	0.8
Hb mmol/l **	3.9	0.3	8.9	0.4	11.9	0.5
Hct %	17.7	1.5	40.3	2.0	53.2	2.4
Hct l/l	0.177	0.015	0.403	0.020	0.532	0.024
MCV fl	81.0	5.0	90.0	5.0	97.0	5.0
MCH pg **	28.8	2.5	32.1	2.5	34.9	2.5
MCH fmol **	1.79	0.16	1.99	0.16	2.16	0.16
MCHC g/dl **	35.5	3.0	35.7	3.0	35.9	3.0
MCHC mmol/l **	22.0	1.8	22.2	1.8	22.3	1.8
RDW-CV	13.8	3.0	13.9	3.0	13.5	3.0
RDW-SD	39.9	6.0	44.7	6.0	46.8	6.0
Plts 10 ⁹ /l	80	20	227	40	436	60
MPV fl	8.1	3.0	8.8	3.0	8.1	3.0
LYMP %	59.0	8.0	30.5	7.0	13.0	6.0
MONO %	11.0	6.0	6.0	6.0	4.5	4.5
GRAN %	30.0	7.0	63.5	7.0	82.5	8.0
LYMP #	1.3	0.2	2.3	0.6	2.4	1.2
MONO #	0.2	0.2	0.5	0.5	0.8	0.8
GRAN #	0.7	0.2	4.9	0.6	15.3	1.6

**** Mindray: BC-3000 with Cymet Mindray (with CN) (old series s/n A59xxx)**

Hb g/dl	6.0	0.4	13.4	0.6	17.8	0.8
Hb mmol/l	3.7	0.3	8.3	0.4	11.0	0.5
MCH pg	27.4	2.5	29.9	2.5	32.5	2.5
MCH fmol	1.70	0.16	1.86	0.16	2.02	0.16
MCHC g/dl	33.8	3.0	33.2	3.0	33.5	3.0
MCHC mmol/l	21.0	1.8	20.6	1.8	20.8	1.8

Mindray BC-2300, BC-2100

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	7.5	1.0	20.0	2.5
RBC 10 ¹² /l	2.32	0.18	4.53	0.24	5.42	0.30
Hb g/dl	6.4	0.4	13.7	0.6	18.3	0.8
Hb mmol/l	4.0	0.3	8.5	0.4	11.4	0.5
Hct %	16.9	1.5	37.8	2.0	50.1	2.4
Hct l/l	0.169	0.015	0.378	0.020	0.501	0.024
MCV fl	73.0	5.0	83.5	5.0	92.5	5.0
MCH pg	27.6	2.5	30.2	2.5	33.8	2.5
MCH fmol	1.71	0.16	1.88	0.16	2.10	0.16
MCHC g/dl	37.8	3.0	36.2	3.0	36.5	3.0
MCHC mmol/l	23.5	1.8	22.5	1.8	22.7	1.8
RDW-CV	13.5	3.0	13.5	3.0	13.0	3.0
RDW-SD	27.0	6.0	30.5	6.0	31.0	8.0
Plts 10 ⁹ /l	71	20	230	40	430	60
MPV fl	11.0	3.0	10.8	3.0	11.0	3.0
PCT %	0.08	0.05	0.25	0.10	0.47	0.20
PCT ml/l	0.78	0.50	2.48	1.00	4.73	2.00
PDW fl	15.0	3.0	14.7	3.0	14.7	3.0
LYMP %	61.0	9.0	31.5	7.0	15.0	6.0
MONO %	7.5	7.0	5.5	5.0	5.5	5.0
GRAN %	31.5	8.0	63.0	7.0	79.5	8.0
LYMP #	1.2	0.2	2.4	0.6	3.0	1.2
MONO #	0.2	0.2	0.4	0.4	1.1	1.1
GRAN #	0.6	0.2	4.7	0.6	15.9	1.6