

REF : 3433-3434-3435-3466-3467-3468-3502-3503-3504

LOT : 218

: 2017-02-19

**Coulter Counter Models (with III Diff. option; for AcT see separate table)**

**MELET MS9** NOTE : Please refer to next page for Erma Values

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.7	0.8	21.0	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.64	0.20	5.71	0.25
Hb g/dl	6.1	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	11.4	0.5
Hct %	17.8	1.5	38.9	2.4	52.6	3.0
Hct l/l	0.178	0.015	0.389	0.024	0.526	0.030
MCV fl	75.1	4.0	83.9	4.0	92.1	4.0
MCH pg	25.9	2.4	29.3	2.8	32.1	2.8
MCH fmol	1.61	0.16	1.82	0.17	1.99	0.19
MCHC g/dl	34.4	3.0	34.9	3.0	34.8	3.0
MCHC mmol/l	21.4	2.0	21.7	2.0	21.6	2.0
RDW %	17.0	3.0	16.1	3.0	15.0	3.0
Pits 10 <sup>9</sup> /l	59	20	248	35	503	60
MPV fl	7.7	2.0	8.1	2.0	8.0	2.0
PDW %	17.2	3.0	16.4	3.0	16.8	3.0
LYMP %	59.2	6.0	32.9	4.0	15.7	4.0
MONO %	11.3	11.3	6.4	6.4	5.0	5.0
GRAN %	29.5	5.0	60.7	5.0	79.3	5.0
LYMP #	1.2	0.2	2.5	0.4	3.3	0.9
MONO #	0.2	0.2	0.5	0.5	1.1	1.1
GRAN #	0.6	0.2	4.7	0.4	16.7	1.1

**ABX Roche: Micros 60, Micros ES 60; Bayer: Advia 60**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.2	0.4	7.3	0.8	19.5	2.0
RBC 10 <sup>12</sup> /l	2.38	0.15	4.60	0.20	5.70	0.25
Hb g/dl	5.8	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.6	0.3	8.4	0.4	11.3	0.5
Hct %	16.6	1.5	37.5	2.4	52.0	3.0
Hct l/l	0.166	0.015	0.375	0.024	0.520	0.030
MCV fl	69.7	4.0	81.5	4.0	91.1	4.0
MCH pg	24.5	2.4	29.5	2.8	32.0	2.8
MCH fmol	1.52	0.16	1.83	0.17	1.99	0.19
MCHC g/dl	35.1	3.0	36.2	3.0	35.1	3.0
MCHC mmol/l	21.8	2.0	22.5	2.0	21.8	2.0
RDW %	15.5	3.0	16.5	3.0	15.4	3.0
Pits 10 <sup>9</sup> /l	64	20	245	35	470	60
MPV fl	7.4	2.0	7.7	2.0	7.5	2.0
LYMP %	51.2	6.0	29.2	4.0	14.3	4.0
MONO %	10.8	10.8	6.3	6.3	3.9	3.9
GRAN %	38.0	5.0	64.5	5.0	81.8	5.0
LYMP #	1.1	0.2	2.1	0.4	2.8	0.9
MONO #	0.2	0.2	0.5	0.5	0.8	0.8
GRAN #	0.8	0.2	4.7	0.4	15.9	1.1

**Abbott Cell-Dyn: 1300, 1500, 1600, 1700, 1800 series**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.3	0.4	7.4	0.8	18.9	2.0
RBC 10 <sup>12</sup> /l	2.40	0.15	4.60	0.20	5.60	0.25
Hb g/dl	6.0	0.4	13.4	0.5	18.1	0.6
Hb mmol/l	3.7	0.3	8.3	0.4	11.2	0.5
Hct %	18.0	1.5	39.0	2.4	52.1	3.0
Hct l/l	0.180	0.015	0.390	0.024	0.521	0.030
MCV fl	75.1	4.0	84.9	4.0	93.1	4.0
MCH pg	25.1	2.4	29.1	2.8	32.4	2.8
MCH fmol	1.56	0.16	1.81	0.17	2.01	0.19
MCHC g/dl	33.4	3.0	34.3	3.0	34.7	3.0
MCHC mmol/l	20.8	2.0	21.3	2.0	21.6	2.0
Pits 10 <sup>9</sup> /l	55	20	243	35	490	60
MPV fl	8.6	2.0	8.6	2.0	8.2	2.0
LYMP %	59.2	6.0	32.9	4.0	15.7	4.0
MID %	11.3	11.3	6.4	6.4	5.0	5.0
GRAN %	29.5	5.0	60.7	5.0	79.3	5.0
LYMP #	1.4	0.2	2.4	0.4	3.0	0.9
MID #	0.3	0.3	0.5	0.5	0.9	0.9
GRAN #	0.7	0.2	4.5	0.4	15.0	1.1

**AVL (Contraves): AVL - 816**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	64.0	6.0	39.2	4.0	18.7	4.0
MONO %	10.4	10.4	7.3	7.3	6.4	6.4
GRAN %	25.6	5.0	53.5	5.0	74.9	5.0
LYMP #	1.3	0.2	3.0	0.4	3.9	0.9
MONO #	0.2	0.2	0.6	0.6	1.3	1.3
GRAN #	0.5	0.2	4.1	0.4	15.7	1.1

**Coulter Counter: AcT diff, AcT diff 2**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.7	0.8	20.9	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.64	0.20	5.71	0.25
Hb g/dl	6.1	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	11.4	0.5
Hct %	17.8	1.5	39.4	2.4	53.2	3.0
Hct l/l	0.178	0.015	0.394	0.024	0.532	0.030
MCV fl	75.1	4.0	84.9	4.0	93.1	4.0
MCH pg	25.9	2.4	29.3	2.8	32.1	2.8
MCH fmol	1.60	0.16	1.82	0.17	1.99	0.19
MCHC g/dl	34.4	3.0	34.5	3.0	34.5	3.0
MCHC mmol/l	21.4	2.0	21.4	2.0	21.4	2.0
RDW %	17.0	3.0	16.1	3.0	15.0	3.0
Pits 10 <sup>9</sup> /l	59	20	246	35	501	60
MPV fl	7.7	2.0	8.1	2.0	8.0	2.0
LYMP %	60.9	6.0	32.5	4.0	14.4	4.0
MONO %	8.5	8.5	4.8	4.8	4.1	4.1
GRAN %	30.6	5.0	62.7	5.0	81.5	5.0
LYMP #	1.3	0.2	2.5	0.4	3.0	0.9
MONO #	0.2	0.2	0.4	0.4	0.9	0.9
GRAN #	0.6	0.2	4.8	0.4	17.0	1.1

**Coulter Counter: AcT 8, AcT 10**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table AcT Diff for CBC parameters						
LYMP %	47.3	6.0	24.4	4.0	9.9	4.0
LYMP #	0.9	0.2	1.9	0.4	2.1	0.8

**Nihon Kohden: Celltac and MEK series (not for 5-part diff)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.4	7.7	0.8	21.0	2.0
RBC 10 <sup>12</sup> /l	2.35	0.15	4.64	0.20	5.62	0.25
Hb g/dl	6.3	0.4	13.9	0.5	19.1	0.6
Hb mmol/l	3.9	0.3	8.6	0.4	11.8	0.5
Hct %	17.7	1.5	39.6	2.4	54.5	3.0
Hct l/l	0.177	0.015	0.396	0.024	0.545	0.030
MCV fl	75.1	4.0	85.3	4.0	97.0	4.0
MCH pg	26.6	2.4	29.9	2.8	33.9	2.8
MCH fmol	1.65	0.16	1.86	0.17	2.10	0.19
MCHC g/dl	35.4	3.0	35.1	3.0	34.9	3.0
MCHC mmol/l	22.0	2.0	21.8	2.0	21.7	2.0
RDW %	15.8	3.0	15.5	3.0	14.2	3.0
Pits 10 <sup>9</sup> /l	60	20	252	35	523	60
MPV fl	7.7	2.0	7.1	2.0	6.5	2.0
LYMP %	63.1	8.0	34.2	4.0	18.4	4.0
MONO %	14.1	14.1	6.9	6.9	4.8	4.8
GRAN %	22.8	8.0	58.9	5.0	76.8	5.0
LYMP #	1.3	0.2	2.6	0.4	3.9	0.9
MONO #	0.3	0.3	0.5	0.5	1.0	1.0
GRAN #	0.5	0.4	4.5	0.4	16.1	1.1

**Danam/Drew Excel18=MWI BT2100, D3; Orphee Mythic18; Abbott Emerald18**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.4	7.7	0.8	20.0	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.64	0.20	5.71	0.25
Hb g/dl	5.9	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.7	0.3	8.4	0.4	11.4	0.5
Hct %	17.3	1.5	39.4	2.4	53.4	3.0
Hct l/l	0.173	0.015	0.394	0.024	0.534	0.030
MCV fl	72.9	4.0	84.9	4.0	93.5	4.0
MCH pg	25.0	2.4	29.3	2.8	32.1	2.8
MCH fmol	1.55	0.16	1.82	0.17	1.99	0.19
MCHC g/dl	34.3	3.0	34.5	3.0	34.3	3.0
MCHC mmol/l	21.3	2.0	21.4	2.0	21.3	2.0
RDW	15.0	3.0	13.5	3.0	12.5	3.0
Pits 10 <sup>9</sup> /l	59	20	231	35	467	60
MPV fl	7.7	2.0	8.1	2.0	8.0	2.0
LYMP %	57.2	6.0	29.7	4.0	14.3	4.0
MONO %	10.30	10.23	3.30	3.24	3.00	2.94
GRAN %	32.5	5.0	67.0	5.0	82.7	5.0
LYMP #	1.1	0.2	2.3	0.4	2.9	0.9
MONO #	0.21	0.14	0.25	0.19	0.60	0.54
GRAN #	0.7	0.2	5.2	0.4	16.5	1.1

**AVL (Contraves): AVL - 818**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
Please refer to Table Coulter Counters for CBC parameters						
LYMP %	71.9	6.0	41.3	4.0	19.7	4.0
MONO %	9.1	9.1	6.8	6.8	6.1	6.1
GRAN %	19.0	5.0	51.9	5.0	74.2	5.0
LYMP #	1.5	0.2	3.2	0.4	4.1	0.9
MONO #	0.2	0.2	0.5	0.5	1.3	1.3
GRAN #	0.4	0.2	4.0	0.4	15.6	1.1



The devices as mentioned in this insert comply with the  
In Vitro Diagnostic Medical Devices Directive 98/79/EC.

LOT : 218

🕒 : 2017-02-19

**Diatron\*\*:** Abacus, Abacus 3, Abacus Junior, Arcus; Erba: Elite 3

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	7.7	0.8	21.4	2.0
RBC 10 <sup>12</sup> /l	2.42	0.15	4.76	0.20	6.00	0.25
Hb g/dl	6.2	0.4	13.8	0.5	18.9	0.6
Hb mmol/l	3.8	0.3	8.6	0.4	11.7	0.5
Hct %	18.5	1.5	41.5	2.4	58.0	3.0
Hct l/l	0.185	0.015	0.415	0.024	0.580	0.030
MCV fl	76.6	5.0	87.2	5.0	96.7	5.0
MCH pg	25.6	2.4	29.0	2.8	31.5	2.8
MCH fmol	1.59	0.16	1.80	0.17	1.95	0.19
MCHC g/dl	33.4	3.0	33.3	3.0	32.5	3.0
MCHC mmol/l	20.7	2.0	20.7	2.0	20.2	2.0
RDW-SD fl	58.2	5.0	60.5	5.0	63.1	5.0
RDW-CV %	16.7	3.0	15.3	3.0	14.4	3.0
Plts 10 <sup>9</sup> /l	59	20	248	35	483	60
MPV fl	7.7	2.0	8.4	2.0	8.4	2.0
PCT %	0.05	0.02	0.21	0.06	0.41	0.10
PCT ml/l	0.45	0.20	2.08	0.60	4.06	1.00
PDW-SD fl	11.7	2.5	12.2	2.5	12.7	2.5
PDW-CV %	35.3	5.0	36.9	5.0	38.4	5.0
LYMP %	55.1	6.0	30.9	4.0	12.5	4.0
MONO %	9.1	9.1	5.9	5.9	6.4	6.4
GRAN %	35.7	6.0	63.3	5.0	81.1	6.0
LYMP #	1.1	0.3	2.4	0.4	2.7	0.9
MONO #	0.3	0.3	0.5	0.5	1.4	1.4
GRAN #	0.6	0.2	4.9	0.4	17.4	1.1

\*\* JT Baker reagents CyMet Abacus+Diluid Abacus=Use Reagent Code 2 in service settings!

**Medonic CA530, CA620**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.4	6.6	0.8	16.6	2.0
RBC 10 <sup>12</sup> /l	2.34	0.15	4.61	0.20	5.63	0.25
Hb g/dl	6.2	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.9	0.3	8.4	0.4	11.4	0.5
Hct %	16.5	1.5	36.8	2.4	49.8	3.0
Hct l/l	0.165	0.015	0.368	0.024	0.498	0.030
MCV fl	70.4	4.0	79.8	4.0	88.4	4.0
MCH pg	26.6	2.4	29.5	2.8	32.5	2.8
MCH fmol	1.65	0.16	1.83	0.17	2.02	0.19
MCHC g/dl	37.8	3.0	37.0	3.0	36.8	3.0
MCHC mmol/l	23.5	2.0	23.0	2.0	22.8	2.0
RDW %	15.1	3.0	17.4	3.0	16.9	3.0
Plts 10 <sup>9</sup> /l	58	20	234	35	444	60
MPV fl	7.2	2.0	7.8	2.0	7.7	2.0
LYMP %	65.9	9.0	37.7	6.0	18.8	4.0
MONO %	9.1	9.1	8.3	8.3	7.7	7.7
GRAN %	25.0	5.0	54.0	9.0	73.5	10.0
LYMP #	1.2	0.2	2.5	0.4	3.1	0.9
MONO #	0.2	0.2	0.5	0.5	1.3	1.3
GRAN #	0.5	0.2	3.6	0.4	12.2	1.1

**Inmesco HM1060; Rayto RT-7600; HTI Micros CC 20 Plus**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.4	7.7	0.8	20.4	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.54	0.20	5.51	0.25
Hb g/dl	6.1	0.4	13.9	0.5	19.4	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	12.1	0.5
Hct %	17.8	1.5	39.4	2.4	53.2	3.0
Hct l/l	0.178	0.015	0.394	0.024	0.532	0.030
MCV fl	77.1	4.0	84.9	4.0	94.1	4.0
MCH pg	25.9	2.4	30.6	2.8	35.2	2.8
MCH fmol	1.60	0.2	1.86	0.2	2.19	0.2
MCHC g/dl	34.4	3.0	35.3	3.0	36.5	3.0
MCHC mmol/l	21.4	2.0	21.4	2.0	22.7	2.0
RDW-CV %	12.7	3.0	11.9	3.0	11.0	3.0
RDW-SD fl	37.9	5.0	40.9	5.0	43.1	5.0
Plts 10 <sup>9</sup> /l	51	20	236	35	501	60
MPV fl	9.6	2.0	9.4	2.0	9.0	2.0
PCT %	0.05	0.02	0.22	0.06	0.45	0.10
PCT ml/l	0.49	0.20	2.21	0.60	4.51	1.00
PDW %	9.0	3.0	12.1	3.0	12.8	3.0
LYMP %	53.9	8.0	28.8	5.0	11.9	4.0
MONO %	14.8	14.8	7.3	7.3	7.0	7.0
GRAN %	31.3	6.0	63.9	8.0	81.1	8.0
LYMP #	1.0	0.2	2.2	0.4	2.4	0.9
MONO #	0.3	0.3	0.6	0.6	1.4	1.4
GRAN #	0.6	0.2	4.9	0.8	16.5	1.1

**Avantor BeneSphera™ H32**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.4	7.7	0.8	20.4	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.54	0.20	5.51	0.25
Hb g/dl	6.1	0.4	13.9	0.5	19.4	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	12.1	0.5
Hct %	17.8	1.5	39.4	2.4	53.2	3.0
Hct l/l	0.178	0.015	0.394	0.024	0.532	0.030
MCV fl	77.1	4.0	84.9	4.0	94.1	4.0
MCH pg	25.9	2.4	30.6	2.8	35.2	2.8
MCH fmol	1.60	0.2	1.86	0.2	2.19	0.2
MCHC g/dl	34.4	3.0	35.3	3.0	36.5	3.0
MCHC mmol/l	21.4	2.0	21.4	2.0	22.7	2.0
RDW-CV %	12.7	3.0	11.9	3.0	11.0	3.0
RDW-SD fl	37.9	5.0	40.9	5.0	43.1	5.0
Plts 10 <sup>9</sup> /l	51	20	236	35	501	60
MPV fl	9.6	2.0	9.4	2.0	9.0	2.0
PCT %	0.05	0.02	0.22	0.06	0.45	0.10
PCT ml/l	0.49	0.20	2.21	0.60	4.51	1.00
PDW %	9.0	3.0	12.1	3.0	12.8	3.0
LYMP %	53.9	8.0	28.8	5.0	11.9	4.0
MONO %	14.8	14.8	7.3	7.3	7.0	7.0
GRAN %	31.3	6.0	63.9	8.0	81.1	8.0
LYMP #	1.0	0.2	2.2	0.4	2.4	0.9
MONO #	0.3	0.3	0.6	0.6	1.4	1.4
GRAN #	0.6	0.2	4.9	0.8	16.5	1.1

**ERMA PCE-170, ERMAX-18, PCE-210**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.4	7.7	0.8	20.5	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.78	0.20	5.56	0.25
Hb g/dl	6.1	0.4	13.6	0.5	18.3	0.6
Hb mmol/l	3.8	0.3	8.4	0.4	11.4	0.5
Hct %	17.8	1.5	41.4	2.4	55.6	3.0
Hct l/l	0.178	0.015	0.414	0.024	0.556	0.030
MCV fl	75.1	4.0	86.6	4.0	99.9	4.0
MCH pg	25.9	2.4	28.5	2.8	32.9	2.8
MCH fmol	1.61	0.16	1.77	0.17	2.04	0.19
MCHC g/dl	34.4	3.0	32.9	3.0	33.0	3.0
MCHC mmol/l	21.4	2.0	20.4	2.0	20.5	2.0
RDW %	17.0	3.0	16.1	3.0	15.0	3.0
Plts 10 <sup>9</sup> /l	69	20	242	35	455	60
MPV fl	7.7	2.0	8.1	2.0	8.0	2.0
PDW %	17.2	3.0	16.4	3.0	16.8	3.0
LYMP %	58.9	9.0	32.9	4.0	15.7	4.0
MONO %	8.0	8.0	6.4	6.4	5.0	5.0
GRAN %	33.1	5.0	60.7	9.0	79.3	10.0
LYMP #	1.1	0.2	2.5	0.4	3.2	0.9
MONO #	0.2	0.2	0.5	0.5	1.0	1.0
GRAN #	0.6	0.2	4.7	0.4	16.3	1.1

**Urit 2900 Plus**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.8	0.4	7.7	0.8	22.4	2.0
RBC 10 <sup>12</sup> /l	2.37	0.15	4.76	0.20	5.91	0.25
Hb g/dl	6.0	0.4	13.3	0.5	17.3	0.6
Hb mmol/l	3.7	0.3	8.3	0.4	10.7	0.5
Hct %	17.2	1.5	37.5	2.4	49.1	3.0
Hct l/l	0.172	0.015	0.375	0.024	0.491	0.030
MCV fl	72.6	4.0	78.9	4.0	83.1	4.0
MCH pg	25.4	2.4	27.9	2.8	29.3	2.8
MCH fmol	1.58	0.2	1.73	0.2	1.82	0.2
MCHC g/dl	35.0	3.0	35.4	3.0	35.3	3.0
MCHC mmol/l	21.7	2.0	22.0	2.0	21.9	2.0
RDW-CV %	15.5	3.0	15.1	3.0	14.7	3.0
RDW-SD fl	23.9	5.0	30.9	5.0	43.1	5.0
Plts 10 <sup>9</sup> /l	58	20	228	35	381	60
MPV fl	16.6	2.0	13.3	2.0	12.4	2.0
PCT %	0.10	0.02	0.30	0.06	0.47	0.10
PCT ml/l	0.96	0.20	3.03	0.60	4.72	1.00
PDW %	13.5	3.0	13.8	3.0	14.3	3.0
LYMP %	58.9	8.0	31.0	5.0	12.7	4.0
MONO %	8.0	8.0	4.1	4.1	4.3	4.3
GRAN %	33.1	6.0	64.9	8.0	83.0	8.0
LYMP #	1.1	0.2	2.4	0.4	2.8	0.9
MONO #	0.1	0.1	0.3	0.3	1.0	1.0
GRAN #	0.6	0.2	5.0	0.8	18.6	1.1

LOT : 218  
: 2017-02-19

**Mindray BC-3200, BC-3000CT**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	8.0	1.0	21.7	2.5
RBC 10 <sup>12</sup> /l	2.44	0.18	4.77	0.24	5.90	0.30
Hb g/dl	6.2	0.4	13.9	0.6	19.2	0.8
Hb mmol/l	3.8	0.3	8.6	0.4	11.9	0.5
Hct %	18.4	1.5	41.0	2.0	56.3	2.4
Hct l/l	0.184	0.015	0.410	0.020	0.563	0.024
MCV fl	75.5	5.0	86.0	5.0	95.5	5.0
MCH pg	25.4	2.5	29.1	2.5	32.5	2.5
MCH fmol	1.58	0.16	1.81	0.17	2.02	0.19
MCHC g/dl	33.7	3.0	33.9	3.0	34.1	3.0
MCHC mmol/l	20.9	2.0	21.0	2.0	21.2	2.0
RDW-CV	15.0	3.0	14.0	3.0	13.0	3.0
RDW-SD	35.0	6.0	38.0	6.0	40.0	8.0
Plts 10 <sup>9</sup> /l	69	20	253	40	500	60
MPV fl	8.0	3.0	7.9	3.0	7.5	3.0
LYMP %	57.0	12.0	29.0	8.0	12.5	6.0
MONO %	11.0	9.0	7.0	5.0	6.0	5.0
GRAN %	32.0	10.0	64.0	8.0	81.5	8.0
LYMP #	1.1	0.3	2.3	0.7	2.7	1.3
MONO #	0.2	0.2	0.6	0.5	1.3	1.1
GRAN #	0.6	0.2	5.1	0.7	17.7	1.8

**Mindray BC-2800, BC-2600, BC-2800Vet\*, BC-2600Vet\* (\*SW version 2.0 or higher)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	7.9	1.0	21.3	2.5
RBC 10 <sup>12</sup> /l	2.43	0.18	4.73	0.24	5.90	0.30
Hb g/dl	6.2	0.4	14.0	0.6	19.4	0.8
Hb mmol/l	3.8	0.3	8.7	0.4	12.0	0.5
Hct %	18.2	1.5	40.2	2.0	55.2	2.4
Hct l/l	0.182	0.015	0.402	0.020	0.552	0.024
MCV fl	75.0	5.0	85.0	5.0	93.5	5.0
MCH pg	25.5	2.5	29.6	2.5	32.9	2.5
MCH fmol	1.58	0.16	1.84	0.17	2.04	0.19
MCHC g/dl	34.0	3.0	34.8	3.0	35.2	3.0
MCHC mmol/l	21.1	2.0	21.6	2.0	21.8	2.0
RDW-CV	16.5	3.0	16.0	3.0	15.5	3.0
RDW-SD	40.0	6.0	46.0	6.0	48.0	8.0
Plts 10 <sup>9</sup> /l	59	20	256	40	508	60
MPV fl	9.5	3.0	10.0	3.0	9.7	3.0
PCT %	0.05	0.05	0.25	0.10	0.48	0.20
PCT ml/l	0.50	0.50	2.50	1.00	4.80	2.00
PDW fl	15.2	3.0	14.9	3.0	14.8	3.0
LYMP %	61.0	10.0	30.5	8.0	14.0	7.0
MONO %	9.0	7.0	6.5	5.0	5.0	4.0
GRAN %	30.0	9.0	63.0	9.0	81.0	8.0
LYMP #	1.2	0.2	2.4	0.7	3.0	1.5
MONO #	0.2	0.2	0.5	0.4	1.1	0.9
GRAN #	0.6	0.2	5.0	0.8	17.3	1.8

**Mindray BC-3600, BC-3300**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.5	7.6	1.0	20.8	2.5
RBC 10 <sup>12</sup> /l	2.53	0.18	4.95	0.24	6.05	0.30
Hb g/dl	6.4	0.4	13.9	0.6	19.1	0.8
Hb mmol/l	4.0	0.3	8.6	0.4	11.9	0.5
Hct %	19.2	1.5	41.6	2.0	55.7	2.4
Hct l/l	0.192	0.015	0.416	0.020	0.557	0.024
MCV fl	76.0	5.0	84.0	5.0	92.0	5.0
MCH pg	25.3	2.5	28.1	2.5	31.6	2.5
MCH fmol	1.57	0.16	1.74	0.17	1.96	0.19
MCHC g/dl	33.3	3.0	33.4	3.0	34.3	3.0
MCHC mmol/l	20.7	2.0	20.7	2.0	21.3	2.0
RDW-CV	15.5	3.0	15.0	3.0	14.5	3.0
RDW-SD	35.0	6.0	44.0	6.0	46.5	8.0
Plts 10 <sup>9</sup> /l	58	20	263	40	536	60
MPV fl	8.4	3.0	9.1	3.0	9.0	3.0
PCT %	0.05	0.05	0.23	0.10	0.46	0.20
PCT ml/l	0.50	0.50	2.30	1.00	4.60	2.00
PDW fl	15.2	3.0	15.6	3.0	15.8	3.0
P-LCC	16.0	10.0	80.0	22.0	152.0	30.0
P-LCR	23.5	9.0	31.0	9.0	29.0	9.0
LYMP %	58.0	9.0	29.5	8.0	13.5	8.0
MONO %	12.5	8.0	8.5	7.0	6.0	5.0
GRAN %	29.5	8.0	62.0	8.0	80.5	9.0
LYMP #	1.1	0.2	2.2	0.6	2.8	1.7
MONO #	0.2	0.2	0.6	0.5	1.2	1.0
GRAN #	0.6	0.2	4.7	0.7	16.7	1.9

**Mindray BC-3000 Plus (s/n A75xxx) , BC-2900, BC-1800, Hemolux**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.0	0.5	8.0	1.0	21.6	2.5
RBC 10 <sup>12</sup> /l	2.50	0.18	4.82	0.24	5.98	0.30
Hb g/dl	6.3	0.4	13.9	0.6	19.0	0.8
Hb mmol/l	3.9	0.3	8.6	0.4	11.8	0.5
Hct %	19.0	1.5	41.5	2.0	56.5	2.4
Hct l/l	0.190	0.015	0.415	0.020	0.565	0.024
MCV fl	76.0	5.0	86.0	5.0	94.5	5.0
MCH pg	25.2	2.5	28.8	2.5	31.8	2.5
MCH fmol	1.56	0.16	1.79	0.17	1.97	0.19
MCHC g/dl	33.2	3.0	33.5	3.0	33.6	3.0
MCHC mmol/l	20.6	2.0	20.8	2.0	20.9	2.0
RDW-CV	16.0	3.0	15.0	3.0	14.5	3.0
RDW-SD	34.5	6.0	38.5	6.0	40.0	8.0
Plts 10 <sup>9</sup> /l	70	20	252	40	512	60
MPV fl	7.7	3.0	8.1	3.0	7.7	3.0
PCT %	0.05	0.05	0.20	0.10	0.40	0.20
PCT ml/l	0.50	0.50	2.00	1.00	4.00	2.00
PDW fl	17.5	3.0	16.4	3.0	16.2	3.0
LYMP %	57.5	12.0	28.5	8.0	12.0	6.0
MONO %	11.5	8.0	8.0	7.0	6.5	5.0
GRAN %	31.0	9.0	63.5	8.0	81.5	8.0
LYMP #	1.2	0.3	2.3	0.7	2.6	1.3
MONO #	0.2	0.2	0.6	0.6	1.4	1.1
GRAN #	0.6	0.2	5.1	0.7	17.6	1.8

**Mindray BC-3000 with CyMet Mindray CN-free (old models s/n A59xxx)**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	2.1	0.5	7.7	1.0	19.8	2.5
RBC 10 <sup>12</sup> /l	2.22	0.18	4.47	0.24	5.65	0.30
Hb g/dl **	6.3	0.4	14.5	0.6	19.7	0.8
Hb mmol/l **	3.9	0.3	9.0	0.4	12.2	0.5
Hct %	18.0	1.5	40.2	2.0	54.9	2.4
Hct l/l	0.180	0.015	0.402	0.020	0.549	0.024
MCV fl	81.1	5.0	89.9	5.0	97.1	5.0
MCH pg **	28.5	2.5	32.4	2.5	34.9	2.5
MCH fmol **	1.77	0.16	2.01	0.17	2.17	0.19
MCHC g/dl **	35.1	3.0	36.1	3.0	35.9	3.0
MCHC mmol/l **	21.8	2.0	22.4	2.0	22.3	2.0
RDW-CV	15.8	3.0	15.4	3.0	14.8	3.0
RDW-SD	45.8	6.0	49.3	6.0	51.2	8.0
Plts 10 <sup>9</sup> /l	66	20	230	40	459	60
MPV fl	8.4	3.0	8.9	3.0	8.5	3.0
LYMP %	59.2	12.0	29.9	8.0	13.2	6.0
MONO %	11.3	11.3	6.4	6.4	5.0	5.0
GRAN %	29.5	9.0	63.7	8.0	81.8	8.0
LYMP #	1.2	0.3	2.3	0.7	2.6	1.3
MONO #	0.2	0.2	0.5	0.5	1.0	1.0
GRAN #	0.6	0.2	4.9	0.7	16.2	1.8

**\*\* Mindray BC-3000 with CyMet Mindray (with CN) (old series s/n A59xxx)**

Hb g/dl	6.0	0.4	13.5	0.6	18.4	0.8
Hb mmol/l	3.7	0.3	8.4	0.4	11.4	0.5
MCH pg	27.2	2.5	30.2	2.5	32.6	2.5
MCH fmol	1.69	0.16	1.87	0.17	2.02	0.19
MCHC g/dl	33.5	3.0	33.6	3.0	33.6	3.0
MCHC mmol/l	20.8	2.0	20.9	2.0	20.8	2.0

**Mindray BC-2300, BC-2100**

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 <sup>9</sup> /l	1.9	0.5	7.8	1.0	21.6	2.5
RBC 10 <sup>12</sup> /l	2.36	0.18	4.65	0.24	5.67	0.30
Hb g/dl	6.4	0.4	13.9	0.6	18.9	0.8
Hb mmol/l	4.0	0.3	8.6	0.4	11.7	0.5
Hct %	17.2	1.5	39.0	2.0	52.4	2.4
Hct l/l	0.172	0.015	0.390	0.020	0.524	0.024
MCV fl	73.0	5.0	83.8	5.0	92.4	5.0
MCH pg	27.0	2.5	29.9	2.5	33.2	2.5
MCH fmol	1.68	0.16	1.86	0.16	2.06	0.16
MCHC g/dl	37.0	3.0	35.7	3.0	36.0	3.0
MCHC mmol/l	23.0	1.9	22.2	1.9	22.3	1.9
RDW-CV	15.0	3.0	14.0	3.0	12.9	3.0
RDW-SD	33.0	8.0	36.9	8.0	38.3	8.0
Plts 10 <sup>9</sup> /l	55	20	241	40	511	60
MPV fl	10.1	3.0	10.2	3.0	10.0	3.0
PCT %	0.06	0.06	0.25	0.10	0.51	0.20
PCT ml/l	0.56	0.56	2.45	1.00	5.08	2.00
PDW fl	14.4	3.0	14.8	3.0	15.7	3.0
LYMP %	58.5	10.0	31.9	8.0	13.9	7.0
MONO %	10.4	10.4	6.8	6.8	7.0	7.0
GRAN %	31.1	9.0	61.3	9.0	79.2	8.0
LYMP #	1.1	0.2	2.5	0.7	3.0	1.5
MONO #	0.2	0.2	0.5	0.5	1.5	1.5
GRAN #	0.6	0.2	4.8	0.7	17.1	1.7