



The devices as mentioned in this insert comply with the
In Vitro Diagnostic Medical Devices Directive 98/79/EC.

LOT : 221

📅 : 2017-11-19

Diatron** : Abacus, Abacus 3, Abacus Junior, Arcus; Erba: Elite 3
Human: HumaCount 30^{TS}, HumaCount 80^{TS}

Avantor BeneSpera™ H32

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.5	7.8	0.8	21.3	2.0
RBC 10 ¹² /l	2.34	0.15	4.79	0.20	5.87	0.25
Hb g/dl	5.9	0.4	13.8	0.5	18.3	0.6
Hb mmol/l	3.7	0.3	8.6	0.4	11.3	0.5
Hct %	18.0	1.5	41.8	2.4	55.8	3.0
Hct l/l	0.180	0.015	0.418	0.024	0.558	0.030
MCV fl	77.1	5.0	87.2	5.0	95.2	5.0
MCH pg	25.2	2.4	28.8	2.8	31.1	2.8
MCH fmol	1.56	0.16	1.79	0.17	1.93	0.19
MCHC g/dl	32.7	3.0	33.1	3.0	32.7	3.0
MCHC mmol/l	20.3	2.0	20.5	2.0	20.3	2.0
RDW-SD fl	56.2	5.0	59.4	5.0	62.1	5.0
RDW-CV %	16.1	3.0	15.0	3.0	14.4	3.0
Plts 10 ⁹ /l	65	20	255	35	483	60
MPV fl	8.3	2.0	8.6	2.0	9.1	2.0
PCT %	0.05	0.02	0.22	0.06	0.44	0.10
PCT ml/l	0.54	0.20	2.20	0.60	4.41	1.00
PDW-SD fl	12.6	2.5	12.5	2.5	13.8	2.5
PDW-CV %	35.3	5.0	36.9	5.0	38.4	5.0
LYMP %	57.3	6.0	32.2	4.0	13.2	4.0
MONO %	7.5	7.5	5.3	5.3	5.7	5.7
GRAN %	35.2	6.0	62.5	5.0	81.1	6.0
LYMP #	1.2	0.3	2.5	0.4	2.8	0.9
MONO #	0.3	0.3	0.4	0.4	1.2	1.2
GRAN #	0.6	0.2	4.9	0.4	17.3	1.1

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.4	7.8	0.8	20.3	2.0
RBC 10 ¹² /l	2.29	0.15	4.57	0.20	5.39	0.25
Hb g/dl	5.8	0.4	13.9	0.5	18.8	0.6
Hb mmol/l	3.6	0.3	8.4	0.4	11.7	0.5
Hct %	17.3	1.5	39.6	2.4	51.2	3.0
Hct l/l	0.173	0.015	0.396	0.024	0.512	0.030
MCV fl	77.6	4.0	84.9	4.0	92.6	4.0
MCH pg	25.5	2.4	30.4	2.8	34.9	2.8
MCH fmol	1.58	0.2	1.85	0.2	2.17	0.2
MCHC g/dl	33.7	3.0	35.1	3.0	36.7	3.0
MCHC mmol/l	20.9	2.0	21.3	2.0	22.8	2.0
RDW-CV %	12.0	3.0	11.6	3.0	11.0	3.0
RDW-SD fl	35.9	5.0	39.8	5.0	42.1	5.0
Plts 10 ⁹ /l	57	20	243	35	501	60
MPV fl	10.2	2.0	9.6	2.0	9.7	2.0
PCT %	0.06	0.02	0.23	0.06	0.49	0.10
PCT ml/l	0.58	0.20	2.33	0.60	4.86	1.00
PDW %	9.0	3.0	12.1	3.0	12.8	3.0
LYMP %	56.2	8.0	30.2	5.0	12.8	4.0
MONO %	12.8	12.8	6.7	6.7	6.4	6.4
GRAN %	31.0	6.0	63.1	8.0	80.8	8.0
LYMP #	1.1	0.2	2.4	0.4	2.6	0.9
MONO #	0.3	0.3	0.5	0.5	1.3	1.3
GRAN #	0.6	0.2	4.9	0.8	16.4	1.1

** JT Baker reagents CyMet Abacus+Diluid Abacus=Use Reagent Code 2 in service settings!

Medonic CA530, CA620

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	1.9	0.4	6.7	0.8	16.5	2.0
RBC 10 ¹² /l	2.26	0.15	4.64	0.20	5.51	0.25
Hb g/dl	5.9	0.4	13.6	0.5	17.7	0.6
Hb mmol/l	3.7	0.3	8.4	0.4	11.0	0.5
Hct %	16.0	1.5	37.0	2.4	47.9	3.0
Hct l/l	0.160	0.015	0.370	0.024	0.479	0.030
MCV fl	70.9	4.0	79.8	4.0	86.9	4.0
MCH pg	26.2	2.4	29.3	2.8	32.2	2.8
MCH fmol	1.63	0.16	1.82	0.17	2.00	0.19
MCHC g/dl	37.0	3.0	36.7	3.0	37.0	3.0
MCHC mmol/l	23.0	2.0	22.8	2.0	23.0	2.0
RDW %	14.4	3.0	17.1	3.0	16.9	3.0
Plts 10 ⁹ /l	64	20	241	35	444	60
MPV fl	7.8	2.0	8.0	2.0	8.4	2.0
LYMP %	68.2	9.0	39.1	6.0	19.7	4.0
MONO %	7.1	7.1	7.7	7.7	7.1	7.1
GRAN %	24.7	5.0	53.2	9.0	73.2	10.0
LYMP #	1.3	0.2	2.6	0.4	3.3	0.9
MONO #	0.1	0.1	0.5	0.5	1.2	1.2
GRAN #	0.5	0.2	3.6	0.4	12.1	1.1

ERMA PCE-170, ERMAX-18, PCE-210

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	1.9	0.4	7.8	0.8	20.4	2.0
RBC 10 ¹² /l	2.29	0.15	4.81	0.20	5.44	0.25
Hb g/dl	5.8	0.4	13.6	0.5	17.7	0.6
Hb mmol/l	3.6	0.3	8.4	0.4	11.0	0.5
Hct %	17.3	1.5	41.6	2.4	53.5	3.0
Hct l/l	0.173	0.015	0.416	0.024	0.535	0.030
MCV fl	75.6	4.0	86.6	4.0	98.4	4.0
MCH pg	25.5	2.4	28.3	2.8	32.6	2.8
MCH fmol	1.58	0.16	1.75	0.17	2.02	0.19
MCHC g/dl	33.7	3.0	32.7	3.0	33.1	3.0
MCHC mmol/l	20.9	2.0	20.3	2.0	20.5	2.0
RDW %	16.3	3.0	15.8	3.0	15.0	3.0
Plts 10 ⁹ /l	75	20	249	35	455	60
MPV fl	8.3	2.0	8.3	2.0	8.7	2.0
PDW %	17.2	3.0	16.4	3.0	16.8	3.0
LYMP %	61.2	9.0	34.3	4.0	16.6	4.0
MONO %	6.0	6.0	5.8	5.8	4.4	4.4
GRAN %	32.8	5.0	59.9	9.0	79.0	10.0
LYMP #	1.2	0.2	2.7	0.4	3.4	0.9
MONO #	0.2	0.2	0.5	0.5	0.9	0.9
GRAN #	0.6	0.2	4.7	0.4	16.1	1.1

Inmesco HM1060; Rayto RT-7600; HTI Micros CC 20 Plus

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	2.0	0.4	7.8	0.8	20.3	2.0
RBC 10 ¹² /l	2.29	0.15	4.57	0.20	5.39	0.25
Hb g/dl	5.8	0.4	13.9	0.5	18.8	0.6
Hb mmol/l	3.6	0.3	8.4	0.4	11.7	0.5
Hct %	17.3	1.5	39.6	2.4	51.2	3.0
Hct l/l	0.173	0.015	0.396	0.024	0.512	0.030
MCV fl	77.6	4.0	84.9	4.0	92.6	4.0
MCH pg	25.5	2.4	30.4	2.8	34.9	2.8
MCH fmol	1.58	0.2	1.85	0.2	2.17	0.2
MCHC g/dl	33.7	3.0	35.1	3.0	36.7	3.0
MCHC mmol/l	20.9	2.0	21.3	2.0	22.8	2.0
RDW-CV %	12.0	3.0	11.6	3.0	11.0	3.0
RDW-SD fl	35.9	5.0	39.8	5.0	42.1	5.0
Plts 10 ⁹ /l	57	20	243	35	501	60
MPV fl	10.2	2.0	9.6	2.0	9.7	2.0
PCT %	0.06	0.02	0.23	0.06	0.49	0.10
PCT ml/l	0.58	0.20	2.33	0.60	4.86	1.00
PDW %	9.0	3.0	12.1	3.0	12.8	3.0
LYMP %	56.2	8.0	30.2	5.0	12.8	4.0
MONO %	12.8	12.8	6.7	6.7	6.4	6.4
GRAN %	31.0	6.0	63.1	8.0	80.8	8.0
LYMP #	1.1	0.2	2.4	0.4	2.6	0.9
MONO #	0.3	0.3	0.5	0.5	1.3	1.3
GRAN #	0.6	0.2	4.9	0.8	16.4	1.1

Urit 2900 Plus

Parameters	LOW		NORMAL		HIGH	
	Mean	Range ±	Mean	Range ±	Mean	Range ±
WBC 10 ⁹ /l	1.9	0.4	7.8	0.8	22.3	2.0
RBC 10 ¹² /l	2.29	0.15	4.79	0.20	5.79	0.25
Hb g/dl	5.7	0.4	13.3	0.5	16.7	0.6
Hb mmol/l	3.6	0.3	8.3	0.4	10.4	0.5
Hct %	16.7	1.5	37.8	2.4	47.3	3.0
Hct l/l	0.167	0.015	0.378	0.024	0.473	0.030
MCV fl	73.1	4.0	78.9	4.0	81.6	4.0
MCH pg	25.0	2.4	27.8	2.8	28.9	2.8
MCH fmol	1.55	0.2	1.72	0.2	1.79	0.2
MCHC g/dl	34.2	3.0	35.2	3.0	35.4	3.0
MCHC mmol/l	21.2	2.0	21.8	2.0	22.0	2.0
RDW-CV %	14.8	3.0	14.8	3.0	14.7	3.0
RDW-SD fl	21.9	5.0	29.8	5.0	42.1	5.0
Plts 10 ⁹ /l	64	20	235	35	381	60
MPV fl	17.2	2.0	13.5	2.0	13.1	2.0
PCT %	0.11	0.02	0.32	0.06	0.50	0.10
PCT ml/l	1.10	0.20	3.17	0.60	4.99	1.00
PDW %	13.5	3.0	13.8	3.0	14.3	3.0
LYMP %	61.2	8.0	32.4	5.0	13.6	4.0
MONO %	6.0	6.0	3.5	3.5	3.7	3.7
GRAN %	32.8	6.0	64.1	8.0	82.7	8.0
LYMP #	1.2	0.2	2.5	0.4	3.0	0.9
MONO #	0.1	0.1	0.3	0.3	0.8	0.8
GRAN #	0.6	0.2	5.0	0.8	18.4	1.1

