



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URS-10 100 Teco** MANUFACTURED DATE: 11/11/2010
Finished Kit
 CATALOG#: URS-10-100
 LOT #: 28719 EXP. DATE: 01/2013

<u>COMPONENTS:</u>	<u>CATALOG#:</u>	<u>LOT #:</u>	<u>EXP. DATE</u>
LEUKOCYTES	URSP-LEU	27571	01/2013
NITRITE	URSP-NIT	28020	02/2013
UROBILINOGEN	URSP-URO	28173	02/2013
PROTEIN	URSP-PRO	28117	02/2013
pH	URSP-PH	27587	01/2013
BLOOD	URSP-BLO	28014	02/2013
SPECIFIC GRAVITY	URSP-SG	28172	02/2013
KETONE	URSP-KET	28234	02/2013
BILIRUBIN	URSP-BIL	28019	02/2013
GLUCOSE	URSP-GLU	28118	02/2013

TOLERANCE LEVELS OF CONTROLS

Controls	<u>Leuko- cytes</u>	<u>Nitrite</u>	<u>Urobili- nogen</u>	<u>Protein</u>	<u>pH</u>	<u>Blood</u>	<u>Specific Gravity</u>	<u>Ketone</u>	<u>Bilirubin</u>	<u>Glucose</u>
KOVA I LOT#:139036-4 EXP: 01/2012	2+ to 3+	POS (+)	4.0 to ≥8.0 EU/dl	100 to ≥300 mg/dl	7.0 to 8.0	2+ to 3+	1.010 to ≥1.030	40 to ≥80 mg/dl	2+ to 3+	500 to ≥1000 mg/dl
KOVA II LOT#:138183-2 EXP: 07/2011	Trace to 2+	POS (+)	0.2 to 1.0 EU/dl	NEG (-) to 30 mg/dl	6.5 to 8.5	Trace to 2+	≤1.005 to 1.015	Trace to ≥40 mg/dl	1+ to 2+	Trace to 500 mg/dl
KOVA III LOT#:139316-9 EXP: 08/2012	NEG (-)	NEG (-)	0.2 to 1.0 EU/dl	NEG (-)	5.0 to 6.5	NEG (-)	≤1.005 to 1.015	NEG (-)	NEG (-)	NEG (-)

TEST VALUES/RESULTS

<u>Controls</u>	<u>Leuko- cytes</u>	<u>Nitrite</u>	<u>Urobili- nogen</u>	<u>Protein</u>	<u>pH</u>	<u>Blood</u>	<u>Specific Gravity</u>	<u>Ketone</u>	<u>Bilirubin</u>	<u>Glucose</u>
KOVA I	3+ 3+	+ +	8.0 8.0	300 300	7.5 7.5	3+ 3+	1.015 1.010	40 40	3+ 2+	1000 1000
KOVA II	2+ 2+	+ +	0.2 1.0	Trace Trace	7.5 7.5	1+ 1+	1.015 1.010	15 15	2+ 1+	250 100
KOVA III	- -	- -	0.2 0.2	- -	5.0 6.0	- -	1.015 1.010	- -	- -	- -

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 11/11/2010