



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URS-1K** MANUFACTURE DATE: 12/22/09
CATALOG #: URS-1K-100
LOT #: 23202 EXP. DATE: 5/12

COMPONENT: URS Paper - Ketone LOT#: 22028 EXP. DATE: 05/2012

TOLERANCE LEVELS OF CONTROLS

CONTROLS	LOT #	EXP. DATE	RANGE
KOVA I	134042	08/2010	40 - \geq 80 mg/dL
KOVA II	133585	05/2010	Trace - \geq 40 mg/dL
KOVA III	134047	05/2011	Negative (-)

TEST VALUES/RESULTS

CONTROLS	1	2	# PASS
KOVA I	80	80	2/2
KOVA II	40	40	2/2
KOVA III	-	-	2/2

COMMENTS: VISUAL CHECK OK.

APPROVED BY:  DATE: 12/22/09
Quality Control