TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URS-1P** MANUFACTURE DATE: 11/09/2009
CATALOG #: URS-1P-100
LOT #: 21666 EXP. DATE: 04/2012

COMPONENT: URS Paper - Protein LOT #: 21057 EXP. DATE: 04/2012

TOLERANCE LEVELS OF CONTROLS

CONTROLS	LOT #	EXP. DATE	PROTEIN
Kova-Trol I	134042	08/2010	100 to ≥300
Kova-Trol II	133585	05/2010	Negative (-) to 30
Kova-Trol III	134047	05/2011	Negative (-)

TEST VALUES/RESULTS

CONTROLS	1	2	# PASS
Level I	300	300	2 / 2
Level II	Trace	Trace	2 / 2
Level III	-	-	2 / 2

COMMENTS: VISUAL CHECK OK.

APPROVED BY: Quality Control DATE: 11/09/2009