



CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URS-1P 100 Teco F/K** MANUFACTURE DATE: 10/06/2010
 CATALOG #: URS-1P-100
 LOT #: 28110 EXP. DATE: 01/2013

COMPONENT: URS Paper - Protein LOT #: 27552 EXP. DATE: 01/2013

TOLERANCE LEVELS OF CONTROLS

CONTROLS	LOT #	EXP. DATE	PROTEIN
Kova-Trol I	139036-4	01/2012	100 to \geq 300
Kova-Trol II	138183-2	07/2011	Negative (-) to 30
Kova-Trol III	137752-3	02/2012	Negative (-)

TEST VALUES/RESULTS

CONTROLS	1	2	# PASS
Level I	300	300	2 / 2
Level II	Trace	Trace	2 / 2
Level III	-	-	2 / 2

COMMENTS: VISUAL CHECK OK.

APPROVED
 BY:

V. Casey
 Quality Control

DATE: 10/07/2010