

TECO DIAGNOSTICS

1268 N. Lakeview Ave., Anaheim, CA 92807, U.S.A.
Phone: (714) 463-1111 Toll Free (USA): (800) 222-9880
Fax: (714) 463-1169
www.tecodiag.com

CERTIFICATE OF ANALYSIS
(FINISHED KIT QUALITY CONTROL)

PRODUCT NAME: **URS-2P 100 Teco Finished Kit** MANUFACTURE DATE: 12/16/2010
CATALOG #: **URS-2P-100**
LOT #: 29729 EXP. DATE: 02/2013

COMPONENTS: URS PAPER – GLUCOSE LOT #: 28118 EXP. DATE: 02/2013
URS PAPER – PROTEIN LOT #: 28117 EXP. DATE: 02/2013

TOLERANCE LEVELS OF CONTROLS

CONTROLS	LOT #	EXP. DATE	GLUCOSE	PROTEIN
KOVA I	139036-4	01/2012	500 to ≥ 2000 mg/dl	100 to ≥ 2000 mg/dl
KOVA II	138183-2	07/2011	Trace to 500 mg/dl	Negative (-) to 30 mg/dl
KOVA III	140029-4	12/2012	Negative (-)	Negative (-)

TEST VALUES/RESULTS

CONTROLS	GLUCOSE		PROTEIN		# PASS
	1	2	1	2	
KOVA I	1000	1000	300	300	4/4
KOVA II	250	250	Trace	Trace	4/4
KOVA III	-	-	-	-	4/4

COMMENTS: VISUAL CHECK OK.

APPROVED
BY:


Quality Control

DATE: 12/16/2010